

Regarding requirement of Hot Air Oven at BPS Govt. Medical College for Women Khanpur Kalan Sonapat. The details of the items are as under:-

Sr. No.	Name of the items	Quantity	Specification
1.	Hot Air Oven	01 (increase & decrease possible)	Attached

All interested firms are invited to submit their sealed quotation rates by 21/10/2024 till 11:00 AM. It is requested to the firms please put Name of the Work, Full Address, Mobile Number of the firm on envelop. The quotation will not be accepted after the due date and time.



Sr. Accounts Officer
For Director BPS GMC for Women
Khanpur Kalan, (Sonapat)



Specifications Hot Air Oven

- Capacity 250 Litre.
- Temp range: +5°C to 300°C.
- Control accuracy: $\pm 0.5^\circ\text{C}$.
- Uniformity: ± 2.0 at 100.0°C
- Minimum 3 No. of Shelves with height adjustable in 25mm steps.
- Glass window in-built into the door for easy viewing of samples.
- Automatic cut off of heater & blower when door opened.
- Digital PID temperature controller with timer, alarms and auto tuning.
- Aero dynamic internal design for achieving horizontal air circulation.
- Solid and plain bottom without electricals.
- Outer body made of G.I Epoxy Coated
- Inner body made of Stainless steel with clear bottom
- CE certified.
- Warranty period (minimum 5 year) to be mentioned.
- CMC should be 5 years.
- Audio / Visual alarm (High, Low temp, Door open)

Signature
Signature

Signature