

**Bhagat Phool Singh Govt. Medical College For Women
Khanpur Kalan, Sonipat.**

Email: bpsgmc.purchase@gmail.com

**Invitation of e-quotation for
CTG Machine**

Inquiry No: Purchase/2025/ **6133**

Inquiry Issue Date: 28.05.2025

Last Date of Submission: 10.06.2025 (12:00 PM)

Rate quotation in sealed envelope/Password protected quotation are hereby invited from the valid firms and suppliers by the undersigned on behalf of the Director, BPSGMC Khanpur Kalan for Supply of Consumables as per Annexure-1 for the Institute as per terms & conditions mentioned below. The filled e-quotations must receive through registered e-mail (Password protected) on or before 10.06.2025 at 12:00 PM. & quotation opened on 10.06.2025 after 01:00 PM. **The interested firm may visit the Institute and quoted the rates thereafter.**

QUOTATION FOR – CTG MACHINE

Closing Date 10.06.2025 Time 12:00 PM

1) Terms & Conditions:-

- a. The quotations received online after this deadline shall not be entertained under any circumstances whatsoever. In case of any delay this Institute will not be responsible.

The online submission through password protected mode will also be restricted.

- b. Quotations must be in online prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative. In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be submitted in the office of undersigned before deadline of submitting the quotation.
- c. Rates must be quoted in **Indian Rupees** and as per the format specified taxes extra if any must be written separately.
- d. Rates must be quoted FOR basis (including Freight charges)
- e. No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- f. The rates quoted must be valid for 60 days minimum or as per State Govt. instruction issued from time to time, from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- g. Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- h. Any conditional quotation shall be rejected summarily.
- i. **Delivery Period** –As per supply order on issuing by this office.
- j. **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value on which amount when there is no EMD.

- k. **Payment Terms:** Payment will be only after satisfactorily complete delivery /commissioning of material and after inspection by Inspection Committee.
- l. BPSGMC, Khanpur Kalan Sonipat reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the BPSGMC, Khanpur Kalan Sonipat will be final in this regard.
- m. BPSGMC, Khanpur Kalan Sonipat reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the BPSGMC, Khanpur Kalan Sonipat will be final in this regard.

2. Special Terms & Conditions:-

- a) Bidder must quote the product as per specification provided in Annexure 1.
- b) The supplier may be asked to submit the sample of quoted make for technical evaluation, to the BPSGMC, Khanpur Kalan Sonipat, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.
- c) Inspection committee will check the products thoroughly, if somehow inspection committee found any discrepancy and is not satisfied with the final product (final product should be same as reviewed ones at the institution) then BPSGMC, Khanpur Kalan, Sonipat has the right to reject the supply and cancel the order, and no claim for payment in this regard will be entertained.
- d) **Jurisdiction:** The Jurisdiction will be district Sonipat only.
- e) It is directed to Incharge Website that requisition of above item may also be placed on website of BPS GMC Khanpur Kalan Sonipat for vide publicity to invite a competitive manner of purchase.


Accounts Officer 

Encl.: Annexure 1 (Specification)

Annexure 2 (Format of price bid)

Annexure 1

S.No.	Name of the item	Qtys	Specification
1	CTG Machine	02	Attached

ANNEXURE "2"

[On the letterhead of firm]
PRICE BIDFORM

To,

Director,
BPSGMC (W),
Khanpur Kalan, Sonipat

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY.....AGAINST THE INQUIRY NO: Purchase/25/.... due on dated.....at BPSGMC, Khapur Kalan, Sonapat.
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. The vendor should also certify that the rates quoted in the quotation are not more than the MRP of the manufacturers.
4. I/We hereby offer to supply at the following rates.

Sr. No	Particular	Quantity	Quoted Make with warranty period (If any)	Price/Unit Exclusive of Tax (INR)	GST/CST/ST
1.					
2.					

Date _____

Place _____

(Signature of Authorized Person)_____

(Name)_____

Phone _____

TECHNICAL SPECIFICATIONS OF FOETAL MONITOR

Sr. No.	Technical Specification	Compliance on each parameter with detailed substantiation how the offered product meets the requirement. (Simply writing as YES/ Complied/ As per BIS/CE/ISO terms is not allowed)	Remarks, if any
1.	The system should be Microprocessor based Foetal Monitor providing continuous monitoring of foetal heart rate (FHR) alongwith maternally sensed foetal activity during antepartum testing for NST (Non-StressTest) and for intensive monitoring of active labor, with twin foetal monitoring facility at the same time.		
2.	Transducer		
i	Type: Multicrystal wide-beam transducer.		
ii	Technique: Autocorrelation.		
iii	Quantity :2 nos (FHR 1, FHR 2)		
iv	Frequency : 1MHz to 2 MHz.		
v	Intensity : Less than 10mW/Sq cm.		
vi	Resolution: 1BPM.		
vii	Heart Rate counting Range: 30 to 250BPM.		
3.	Printer : Facility to print on Inbuilt thermal printer (On thermal paper) as well as on plain paper via any Deskjet/Laser printer (Should provide connectivity for Deskjet/Laser Printer).		
i	Paper : Z-fold Pre-printed chart scale.		
ii	Speeds: 1 to 3cm/ minute.		
4.	Features:		
i	1. Twin fetal monitoring with TOCO transducers – 01 No.		
ii	2. It should have clinical event marker.		
iii	3. It should have monitoring of Bradycardia & Tachycardia alarm events.		

	4. It should have facility to control the volume of FHR sound.		
v	5. It should have battery back up of 4-6 hours.		
vi	6. Power Supply: 230Vac, 50/60Hz.		
5	Accessories :		
i	1. Should provide rechargeable battery along with recharging unit (Charger/ Adaptor).		
ii	2. Should provide a pre-cut non-fray elasticized belt with buckle shall enable easy transducer positioning for more accurate traces.- 03Nos.		
iii	3. Vibroacoustic stimulator – 01 No.		
6	Display:		
i	1. Display Minimum 5".		
ii	2. Actual FHR1 & FHR2 in BPM.		
iii	3. Uterine Contraction/Activity in %.		
iv	4. High / Low FHR limits.		
v	5. Alarm Message Display.		
vi	6. Battery charging and Low indication.		
vii	7. Blinking corresponding to each beat		
7	Patient Database: It should store more than 10 hours, memory to store all the data. In case the printer goes out of order, machine should continue monitoring and then download to printer afterwards for printing or be viewed.		
8	Computer interface through RS232 connection.		
9	Unit quoted should have unique model number.		
10	Units supplied should hold a unique Serial/Identification number.		
11	Two years comprehensive warranty.		
12	Should be CE (Should be from Notified Body) or US FDA Approved		
13	Should quote Standard Accessories separately.		
14	Should Provide Extra Ultrasound Transducer- 01 No. & TOCO Transducer – 01 No., pre-cut non-fray elasticized belt- 17Nos. - 1		
15	Demonstration is compulsory		
16	Should Provide Training to the end user and Biomedical Engineer.		
17	Tropicalisation:		

	1. Operating room temp. 40deg. C		
	2. Storage room temp. 60deg. C		
	3. Relative Humidity 90% Non-condensing		
18	Should quote Cost of AMC and CMC separately.		
19	Should Provide Service, Spare parts and consumables support for Ten years.		

1
Checked & verified

h
7/12/24.