


Regarding requirement of Stripper for Blood Tubing at BPS Govt. Medical College for Women Khanpur Kalan Sonapat. The details of the items are as under:-

<b>Sr. No.</b>	<b>Name of the items</b>	<b>Quantity</b>	<b>Specification</b>
1.	Stripper for Blood Tubing	02 (increase & decrease possible)	Attached

All interested firms are invited to submit their sealed quotation rates by 21/10/2024 till 11:00 AM. It is requested to the firms please put Name of the Work, Full Address, Mobile Number of the firm on envelop. The quotation will not be accepted after the due date and time.

  
Sr. Accounts Officer  
For Director BPS GMC for Women  
Khanpur Kalan, (Sonapat)

## Technical Specification – Stripper

1. Completely stainless steel body, so no chance of rusting ever.
2. Light weight
3. Spring powered handle, to ensure uniform pressure while pressing to close and automatic recoiling of spring to release handle for opening.
4. Screw – less rollers to avoid loosening of the rollers.
5. Extra sharp cutting edges.
6. Hot melt plastic coated handle covers to ensure no space or gap in between the handle arms and plastic covers.
7. Ergonomically designed handle for better gripping.
8. Roller guide to avoid any damage of tube.
9. Manual tube sealing by aluminium rings.
10. Warranty for 5 years, <sup>CMC/</sup>AMC for 5 years.

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*  
Singh Rajiv

*[Handwritten signature]*